



In a cup-sized sample, dip strip for 2 seconds and remove. Hold strip level (pad face up). Wait 10 seconds. Compare to color chart.

Low Range TOTAL Chlorine ppm					
0	0.1	0.25	0.5	1	3 10

TIPS

- Keep wet fingers out of the vial.
- Close vial tightly after removing strip.
- Store in a cool, dry place.

Code 2963LR-G
 Lot # 8367 Exp. 7/10